

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

SEP 10 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130571</u>	
Name <u>JSwartz</u>		ID# <u>4314</u>	Date <u>8/4/09</u>
A	Agency <u>SCDL</u>	Phone # <u>269-5592</u>	
Instrument Location <u>SCDL</u>			
Alco S/N <u>SC01108</u>		Target Value <u>.080</u>	High Pressure <u>700psi</u>
B	Alco Test Values		
	<u>.078</u> 1 st Alco	<u>.080</u> 2 nd Alco	
Signature <u>[Signature]</u>			
(OVER)			

BMB
9/10/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 15th day of Oct, 2009.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

AUGUST 04, 2009 ✓ *amb 9/10/09*

OPERATOR'S NAME:

SWARTZ/J

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: AAA1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .000 ✓

ALCO S/N: SCOTT08 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.03 in
ALCO TARGET .000 17:38
BLANK TEST .000 17:38
INTERNAL STANDARD VERIFIED 17:38
ALCO TV 30.03 in .078 17:39
BLANK TEST .000 17:39
SUBJECT SAMPLE *amb 9/10/09* .000 17:40
BLANK TEST .000 17:41
ALCO TV 30.03 in .000 17:41
BLANK TEST .000 17:42

amb 08/04/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

AUGUST 04, 2009 ✓
TIME 17:36

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 12/11/00
HEATERS
SAMPLE CHAMBER: 51c
BREATH TUBE: 40c
BAROMETER: 30.03 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`ab cde fgh i j k l m n o
p q r s t u v w x y z { | } ~ "